



90XL Meter Time Keeping Certification

Mesa Laboratories, Inc. certifies that all products it designs and sells have been extensively validated to assure that they meet or exceed all design, performance and accuracy specifications. The quartz crystal employed in 90XL for its time keeping function is specified at better than +/- 30 ppm accuracy, resulting in an overall accuracy of better than 1 sec./hr. The manufacturing test, and calibration processes and procedures utilized in manufacturing and servicing the 90XL assures that each meter keeps accurate time.